

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和7年3月3日

経理責任者

独立行政法人国立病院機構

本部総務部長 坊野 静成

◎調達機関番号 597 ◎所在地番号 13

1 調達内容

- (1) 品目分類番号 22、31
- (2) 導入計画物品及び数量
 - ①マルチスライスCT 16式
 - ②MRI 14式
 - ③血管連続撮影装置 19式
 - ④ガンマカメラ 14式
 - ⑤高度放射線治療装置 3式
 - ⑥X線透視撮影装置 21式
 - ⑦X線発生装置 17式
 - ⑧FPDシステム 15式
 - ⑨乳房X線撮影装置 6式
 - ⑩外科用X線撮影装置 23式

2 調達方法 購入（国立病院機構、労働者健康安全機構、地域医療機能推進機構、日本赤十字社及び国立研究開発法人国立精神・神経医療研究センターにおける共同入札）

3 導入予定時期 ①～⑩ 令和8年3月末

4 調達に必要なとされる基本的な要求要件

- ① マルチスライスCT
 - ・ 1スキャンで16スライス以上撮影することができること。
 - ・ 全肺や全腹部を数秒で撮影する能力があること。
- ② MRI
 - ・ 1.5テスラ以上の磁場強度を有する磁気共鳴画像装置であること。
 - ・ 頭部、躯幹部、四肢などすべての検査で高精細な画像が迅速に撮影でき、解析機能が充実していること。
 - ・ 高速処理可能な高性能な操作コンソールを有し、操作性に優れていること。解析診断用のワークステーションを備えていること。
- ③ 血管連続撮影装置
 - ・ 不整脈の診断、治療と頭腹部及び下肢の血管造影による診断、治療および全身の血管造影による診断、治療が安全に行えること。
 - ・ Cアームは1方向又は2方向を有し、迅速にアームの角度を設定でき、

任意の方向からデジタル透視、デジタル撮影ができること。

- ・フラットパネルディテクタは、血管造影による診断・治療が行えること。
- ・デジタル画像処理システムは血管内治療を支援する機能や解析機能を有すること。

④ガンマカメラ

- ・多検出器デジタルガンマカメラ装置であること。
- ・スペクト収集とホールボディ収集の機能を有していること。
- ・フュージョンソフトや定量性を行えるアプリケーションを十分に備えていること。

⑤高度放射線治療装置

- ・多目的な治療に対応が可能であること。
- ・遮蔽能力の高い多分割コリメータを有すること。
- ・システムとして安全で簡便な操作性を有し、高精度な放射線治療への対応が可能であること。

⑥X線透視撮影装置

- ・透視撮影ができること。
- ・動画記録が可能であること。
- ・寝台の水平位が可能であること。
- ・フラットパネルディテクタを有すること。

⑦X線発生装置

- ・X線管球を有すること。
- ・立位、臥位の撮影台を有し、全身の撮影が迅速かつ安全な撮影への対応が可能であること。

⑧FPDシステム

- ・フラットパネルディテクタを有すること。

⑨乳房X線撮影装置

- ・デジタル乳房X線撮影装置であること。
- ・乳房バイオプシができるシステムを有すること。
- ・乳房撮影診断用高精細モニタシステムを有すること。

⑩外科用X線撮影装置

- ・デジタル外科用X線撮影装置であること。

5 資料及びコメントの提供方法 上記1(2)の物品に関する一般的な参考資料及び4の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提供期限 令和7年4月18日17時00分

(2) 提出先 〒152-8621 東京都目黒区東が丘2-5-21 独立行政法人
国立病院機構本部支援部調達課共同調達係 瀧口 美徳 電話 03-5712-5065

6 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付

する。

(1) 交付期間 令和7年3月3日から令和7年4月18日まで。

(2) 交付場所 上記5(2)に同じ。

7 その他 この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

8 Summary

(1) Classification of the products to be procured: 22,31

(2) Nature and quantity of the products to be purchased :

① Multi-slice Computed Tomography System 16 sets

② Magnetic Resonance Imaging System 14 sets

③ Serial Angiographic System 19 sets

④ Gamma Camera System 14 sets

⑤ High Radiation Therapy System 3 sets

⑥ X-ray fluoroscope Radiography System 21 sets

⑦ X-ray generator System 17 sets

⑧ Flat-panel detector System 15 sets

⑨ Digital X rays equipment to mammography 6 sets

⑩ Mobile X-ray Surgical Imaging System 23 sets

(3) Type of the procurement : Purchase

(4) Basic requirements of the procurement :

① Multi-slice Computed Tomography System

- The system should be capable of scanning more than 16 slices data during one rotation.

- The system should have ability that could scan either whole lung or abdomen in few seconds.

② Magnetic Resonance Imaging System

- Magnetic field strength should be 1.5 Tesla or more.

- The images of examination in head, body and extremities are enable to examine speedily with high resolution images.

- The console unit must provide high performance capability and exceptional in operation.

③ Serial Angiographic System

- The system should be designed safely for interventional procedures in arrhythmia, also in head, abdomen, lower extremities and whole body for interventional angiography.

- C-arm should equip either one or two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary-direction.

- The flat-panel detector should be especially designed to fulf-

ill the interventional angiography.

- The digital imaging processing system should have supporting and analysis function for intravascular treatment.

④ Gamma Camera System

- The system should be multiple digital gamma camera.
- The system should have function of SPECT and Whole body acquisition.
- The system should equip the fusion software and quantitative application replete.

⑤ High Radiation Therapy System

- The system should be capable to multipurpose treatment.
- The system should equip multi-leaf collimator with high shielding performance.
- The system should be designed safely and user-friendliness operability, in addition, conformable to high-precision radiation therapy.

⑥ X-ray fluoroscope Radiography System

- The system is available in fluoroscopic radiography.
- The system should be capable of recording dynamic images.
- The system should be capable of positioning the bed position to horizontal.
- The system should equip the flat-panel detector.

⑦ X-ray generator System

- The system should equip x-ray tube.
- The system should equip a table for standing and lying position, in addition, radiography of whole body could be able to perform speedily and safely.

⑧ Flat-panel detector System

- The system should equip the flat-panel detector.

⑨ Digital X rays equipment to mammography

- The system should be digital X ray-s equipment to mammography.
- The system must be able to digital stereotactic mammography.
- The system should equip the high resolution monitor system for mammography diagnosis.

⑩ Mobile X-ray Surgical Imaging System

- The system should be digital mobile X-ray surgical imaging system.

(5) Time-limit for the submission of the requested material : 5:00 P.M. April 18, 2025

(6) Contact point for the notice : Yoshinori Takiguchi, Chief Clerk,

Supply Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2-5-21 Higashigaoka Meguro-ku Tokyo 152-8621, Japan. TEL03-5712-5065